



**QPS Evaluation Services Inc**  
 Testing, Certification and Field Evaluation Body  
 Accredited in Canada, the USA, and Internationally

File <b>LR1421</b>
-----------------------

**CERTIFICATE OF COMPLIANCE**  
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	OTIS Instruments Inc.
Address	301 S. Texas Avenue Bryan, Texas 77802 USA
Project Number	LR1421-2
Product	Explosion Proof Sensor Housing
Model Number	OI-254
Ratings/Markings	Class I, Division 1, Groups C and D T <sub>Amb</sub> = -40 °C to +50 °C 8.5 Vdc Max., 500mA Max.
Applicable Standards	<ul style="list-style-type: none"> <li>• CSA C22.2 No. 0-10 (R2015)</li> <li>• CSA C22.2 No. 30-M1986(2016)</li> <li>• CSA C22.2 No. 142-M1987(2014)</li> <li>• UL 508 18<sup>th</sup> Edition</li> <li>• UL 1203 5<sup>th</sup> Edition</li> </ul>
Schedule of Limitations	This is a component and is subject to further investigation at the end use.
Factory/Manufacturing Location	Same as Applicant

**Statement of Compliance:** The products identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standards. As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: **Dave Adams, P. Eng.**  
 Manager, Hazardous Locations [Ex Equipment] Department

Signature: 

Date: June 8, 2018